

*Read full Reveal 2.0 for DSP (Neogen item 9561) kit instructions carefully  
before starting test*

*Read full Reveal 2.0 for DSP (Neogen item 9561) kit instructions carefully  
before starting test*

# Reveal<sup>®</sup> 2.0

## DSP Hydrolysis Pack

# Reveal<sup>®</sup> 2.0

## DSP Hydrolysis Pack

### INTENDED USE

DSP Hydrolysis Pack is intended to be used as an accessory to the Reveal 2.0 for DSP test kit for the extraction of OA and DTXs in shellfish samples.

### MATERIALS PROVIDED

DSP Hydrolysis Pack (Neogen item 9554)

1. 2.5 M sodium hydroxide (NaOH) solution (5 mL)
2. 2.5 M hydrochloric acid (HCl) solution (5 mL)
3. 24 Glass vials and caps

### STORAGE

DSP Hydrolysis Pack may be stored at room temperature but should not be frozen or exposed to high temperatures (> 90°F) for more than two weeks. For best results, do not store in direct sunlight.

### DISPOSAL

Dispose in accordance with all applicable federal, state and local environmental regulations.

### CUSTOMER SERVICE

Neogen Customer Assistance and Technical Services can be reached by using the contact information on the back of this booklet. Training on this product, and all Neogen test kits, is available.

### INTENDED USE

DSP Hydrolysis Pack is intended to be used as an accessory to the Reveal 2.0 for DSP test kit for the extraction of OA and DTXs in shellfish samples.

### MATERIALS PROVIDED

DSP Hydrolysis Pack (Neogen item 9554)

1. 2.5 M sodium hydroxide (NaOH) solution (5 mL)
2. 2.5 M hydrochloric acid (HCl) solution (5 mL)
3. 24 Glass vials and caps

### STORAGE

DSP Hydrolysis Pack may be stored at room temperature but should not be frozen or exposed to high temperatures (> 90°F) for more than two weeks. For best results, do not store in direct sunlight.

### DISPOSAL

Dispose in accordance with all applicable federal, state and local environmental regulations.

### CUSTOMER SERVICE

Neogen Customer Assistance and Technical Services can be reached by using the contact information on the back of this booklet. Training on this product, and all Neogen test kits, is available.

### MSDS INFORMATION AVAILABLE

Material safety data sheets (MSDS) are available for this test kit, and all of Neogen's test kits, on Neogen's website at [www.neogen.com](http://www.neogen.com), or by calling Neogen at 800/234-5333 or 517/372-9200.

### WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. If any materials are defective, Neogen will provide a replacement of the product. Buyer assumes all risk and liability resulting from the use of this product. There is no warranty of merchantability of this product or of the fitness of the product for any purpose. Neogen shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product.



#### North America Neogen Headquarters

620 Leshler Place, Lansing, MI 48912 USA  
800/234-5333 (USA/Canada) or 517/372-9200  
Fax: 517/372-2006 • [foodsafety@neogen.com](mailto:foodsafety@neogen.com)  
[www.neogen.com](http://www.neogen.com)

#### Europe, Middle East and Africa Neogen Europe

The Dairy School, Auchincruive, Ayr  
KA6 5HW Scotland, UK  
+ 44 (0) 1292 525 600  
Fax: + 44 (0) 1292 525 601  
[info\\_uk@neogeneurope.com](mailto:info_uk@neogeneurope.com)  
[www.neogeneurope.com](http://www.neogeneurope.com)

#### Mexico Neogen Latinoamérica

Darwin No. 83, Col. Anzures, México, 11590 D.F.  
+52 (55) 5254-8235, +52 (55) 5203-0111,  
+52 (55) 5531-2837  
Fax: +52 (55) 5531-1647  
[informacion@neogenlac.com](mailto:informacion@neogenlac.com) • [www.neogen.com](http://www.neogen.com)

#### Brazil

#### Neogen do Brasil

Rua: Alberto Guizo 760, Distrito Industrial João Narezzi, Indaiatuba – SP Brasil, Cep: 13.347-402  
Tel: +55 19 3935.3727  
[info@neogendobrasil.com.br](mailto:info@neogendobrasil.com.br) • [www.neogen.com](http://www.neogen.com)

### MSDS INFORMATION AVAILABLE

Material safety data sheets (MSDS) are available for this test kit, and all of Neogen's test kits, on Neogen's website at [www.neogen.com](http://www.neogen.com), or by calling Neogen at 800/234-5333 or 517/372-9200.

### WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. If any materials are defective, Neogen will provide a replacement of the product. Buyer assumes all risk and liability resulting from the use of this product. There is no warranty of merchantability of this product or of the fitness of the product for any purpose. Neogen shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product.



#### North America Neogen Headquarters

620 Leshler Place, Lansing, MI 48912 USA  
800/234-5333 (USA/Canada) or 517/372-9200  
Fax: 517/372-2006 • [foodsafety@neogen.com](mailto:foodsafety@neogen.com)  
[www.neogen.com](http://www.neogen.com)

#### Europe, Middle East and Africa Neogen Europe

The Dairy School, Auchincruive, Ayr  
KA6 5HW Scotland, UK  
+ 44 (0) 1292 525 600  
Fax: + 44 (0) 1292 525 601  
[info\\_uk@neogeneurope.com](mailto:info_uk@neogeneurope.com)  
[www.neogeneurope.com](http://www.neogeneurope.com)

#### Mexico Neogen Latinoamérica

Darwin No. 83, Col. Anzures, México, 11590 D.F.  
+52 (55) 5254-8235, +52 (55) 5203-0111,  
+52 (55) 5531-2837  
Fax: +52 (55) 5531-1647  
[informacion@neogenlac.com](mailto:informacion@neogenlac.com) • [www.neogen.com](http://www.neogen.com)

#### Brazil

#### Neogen do Brasil

Rua: Alberto Guizo 760, Distrito Industrial João Narezzi, Indaiatuba – SP Brasil, Cep: 13.347-402  
Tel: +55 19 3935.3727  
[info@neogendobrasil.com.br](mailto:info@neogendobrasil.com.br) • [www.neogen.com](http://www.neogen.com)