

For Laboratory Use

Store at 2–8°C



### ***Pseudomonas Supplement Part #S2-PSI***

(For Isolation of *Pseudomonas aeruginosa* with the PD-109 Vial) • 10 mL per vial

#### **Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-PSI, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 10 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-PSI supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has completely dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

#### **Usage Instructions:**

1. Add 0.1 mL of the rehydrated S2-PSI supplement to the desired *Pseudomonas* test vial.
2. Place the remaining S2-PSI rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120351C

S2-PSI-0819

For Laboratory Use

Store at 2–8°C



### ***Pseudomonas Supplement Part #S2-PSI***

(For Isolation of *Pseudomonas aeruginosa* with the PD-109 Vial) • 10 mL per vial

#### **Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-PSI, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 10 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-PSI supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has completely dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

#### **Usage Instructions:**

1. Add 0.1 mL of the rehydrated S2-PSI supplement to the desired *Pseudomonas* test vial.
2. Place the remaining S2-PSI rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120351C

S2-PSI-0819

For Laboratory Use

Store at 2–8°C



### ***Pseudomonas Supplement Part #S2-PSI***

(For Isolation of *Pseudomonas aeruginosa* with the PD-109 Vial) • 10 mL per vial

#### **Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-PSI, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 10 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-PSI supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has completely dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

#### **Usage Instructions:**

1. Add 0.1 mL of the rehydrated S2-PSI supplement to the desired *Pseudomonas* test vial.
2. Place the remaining S2-PSI rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120351C

S2-PSI-0819

For Laboratory Use

Store at 2–8°C



### ***Pseudomonas Supplement Part #S2-PSI***

(For Isolation of *Pseudomonas aeruginosa* with the PD-109 Vial) • 10 mL per vial

#### **Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-PSI, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 10 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-PSI supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has completely dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

#### **Usage Instructions:**

1. Add 0.1 mL of the rehydrated S2-PSI supplement to the desired *Pseudomonas* test vial.
2. Place the remaining S2-PSI rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120351C

S2-PSI-0819