

BetaStar[®] Combo

BETASTAR[®] Combo Control

Penicillin/Oxytetracycline Positive Control

Contains 25 positive control vials

For use with BetaStar Combo for detection of
beta-lactam and tetracycline residues in raw, commingled cow milk

This positive control preparation is designed to meet requirements for a positive control to be performed daily before milk sample testing and with the initial use of each lot of BetaStar Combo reagents. The positive control consists of penicillin G and oxytetracycline lyophilized in an excipient matrix. When reconstituted with raw milk, the resulting preparation contains penicillin G at a concentration of 4 ± 0.4 ppb, and oxytetracycline at a concentration of 100 ± 10 ppb.

DIRECTIONS FOR USE

1. Remove 1 vial from the box and gently tap the vial against a hard surface to bring all the solid material to the bottom of the vial.
2. Carefully remove the cap and rubber stopper from the vial.
3. Using a fixed volume pipettor, add 1 mL (± 0.05 mL) raw milk* to the positive control vial.
4. Replace stopper and mix the control by inverting the vial.

* Raw milk must be commingled cow milk that previously tested negative on the BetaStar Combo test.

STORAGE REQUIREMENTS

Store lyophilized vials at 2–8°C.

BetaStar[®] Combo

BETASTAR[®] Combo Control

Penicillin/Oxytetracycline Positive Control

Contains 25 positive control vials

For use with BetaStar Combo for detection of
beta-lactam and tetracycline residues in raw, commingled cow milk

This positive control preparation is designed to meet requirements for a positive control to be performed daily before milk sample testing and with the initial use of each lot of BetaStar Combo reagents. The positive control consists of penicillin G and oxytetracycline lyophilized in an excipient matrix. When reconstituted with raw milk, the resulting preparation contains penicillin G at a concentration of 4 ± 0.4 ppb, and oxytetracycline at a concentration of 100 ± 10 ppb.

DIRECTIONS FOR USE

1. Remove 1 vial from the box and gently tap the vial against a hard surface to bring all the solid material to the bottom of the vial.
2. Carefully remove the cap and rubber stopper from the vial.
3. Using a fixed volume pipettor, add 1 mL (± 0.05 mL) raw milk* to the positive control vial.
4. Replace stopper and mix the control by inverting the vial.

* Raw milk must be commingled cow milk that previously tested negative on the BetaStar Combo test.

STORAGE REQUIREMENTS

Store lyophilized vials at 2–8°C.

CUSTOMER SERVICE

Neogen Customer Assistance and Technical Services can be reached by using the contact information on the back of this booklet. Training on this product, and all Neogen test kits, is available.

MSDS INFORMATION AVAILABLE

Material safety data sheets (MSDS) are available for this test kit, and all of Neogen's test kits, on Neogen's website at www.neogen.com, or by calling Neogen at 800/234-5333 or 517/372-9200.

WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. If any materials are defective, Neogen will provide a replacement of the product. Buyer assumes all risk and liability resulting from the use of this product. There is no warranty of merchantability of this product or of the fitness of the product for any purpose. Neogen shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product.



North America

Neogen Headquarters

620 Leshler Place, Lansing, MI 48912 USA
800/234-5333 (USA/Canada) or 517/372-9200
Fax: 517/372-2006 • foodsafety@neogen.com
www.neogen.com

Europe, Middle East and Africa

Neogen Europe

The Dairy School, Auchincruive, Ayr
KA6 5HW Scotland, UK
+ 44 (0) 1292 525 600 • Fax: + 44 (0) 1292 525 601
info_uk@neogeneurope.com
www.neogeneurope.com

Mexico

Neogen Latinoamérica

Darwin No. 83, Col. Anzures, México, 11590 D.F.
(55) 5254-8235, (55) 5203-0111, (55) 5531-2837
Fax: (55) 5531-1647 • informacion@neogenlac.com
www.neogen.com

Brazil

Neogen do Brasil

Rua: Alberto Guizo 760, Distrito Industrial João
Narezzi, Indaiatuba – SP Brasil, Cep: 13.347-402
Tel: +55 19 3935.3727
info@neogendobrasil.com.br • www.neogen.com

CUSTOMER SERVICE

Neogen Customer Assistance and Technical Services can be reached by using the contact information on the back of this booklet. Training on this product, and all Neogen test kits, is available.

MSDS INFORMATION AVAILABLE

Material safety data sheets (MSDS) are available for this test kit, and all of Neogen's test kits, on Neogen's website at www.neogen.com, or by calling Neogen at 800/234-5333 or 517/372-9200.

WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. If any materials are defective, Neogen will provide a replacement of the product. Buyer assumes all risk and liability resulting from the use of this product. There is no warranty of merchantability of this product or of the fitness of the product for any purpose. Neogen shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from the use of this product.



North America

Neogen Headquarters

620 Leshler Place, Lansing, MI 48912 USA
800/234-5333 (USA/Canada) or 517/372-9200
Fax: 517/372-2006 • foodsafety@neogen.com
www.neogen.com

Europe, Middle East and Africa

Neogen Europe

The Dairy School, Auchincruive, Ayr
KA6 5HW Scotland, UK
+ 44 (0) 1292 525 600 • Fax: + 44 (0) 1292 525 601
info_uk@neogeneurope.com
www.neogeneurope.com

Mexico

Neogen Latinoamérica

Darwin No. 83, Col. Anzures, México, 11590 D.F.
(55) 5254-8235, (55) 5203-0111, (55) 5531-2837
Fax: (55) 5531-1647 • informacion@neogenlac.com
www.neogen.com

Brazil

Neogen do Brasil

Rua: Alberto Guizo 760, Distrito Industrial João
Narezzi, Indaiatuba – SP Brasil, Cep: 13.347-402
Tel: +55 19 3935.3727
info@neogendobrasil.com.br • www.neogen.com