

For Laboratory Use

Store at 2–8°C



**Staphylococcus aureus Supplement Part #S2-SSA**

20 mL per vial

**Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-SSA, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 20 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-SSA supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

**Usage Instructions:**

1. Mix the supplement by pipetting up and down quickly.
2. Add 0.2 mL of the rehydrated S2-SSA supplement to the desired Staphylococcus test vial.
3. Place the remaining S2-SSA rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120352B

S2-SSA-0819

For Laboratory Use

Store at 2–8°C



**Staphylococcus aureus Supplement Part #S2-SSA**

20 mL per vial

**Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-SSA, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 20 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-SSA supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

**Usage Instructions:**

1. Mix the supplement by pipetting up and down quickly.
2. Add 0.2 mL of the rehydrated S2-SSA supplement to the desired Staphylococcus test vial.
3. Place the remaining S2-SSA rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120352B

S2-SSA-0819

For Laboratory Use

Store at 2–8°C



**Staphylococcus aureus Supplement Part #S2-SSA**

20 mL per vial

**Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-SSA, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 20 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-SSA supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

**Usage Instructions:**

1. Mix the supplement by pipetting up and down quickly.
2. Add 0.2 mL of the rehydrated S2-SSA supplement to the desired Staphylococcus test vial.
3. Place the remaining S2-SSA rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120352B

S2-SSA-0819

For Laboratory Use

Store at 2–8°C



**Staphylococcus aureus Supplement Part #S2-SSA**

20 mL per vial

**Rehydration Instructions:**

1. Remove the metal cap and sealing ring.
2. Wearing gloves to protect hands and prevent contamination of the S2-SSA, slowly remove the rubber cap and place it top-down on the bench/counter.
3. Add 20 mL sterile Butterfield's Buffer to the vial containing the lyophilized S2-SSA supplement.
4. Replace the rubber cap and gently swirl the vial until the inhibitor has dissolved.
5. Record the date rehydrated on the vial label. Rehydrated supplement can be used for up to four weeks after rehydration.

**Usage Instructions:**

1. Mix the supplement by pipetting up and down quickly.
2. Add 0.2 mL of the rehydrated S2-SSA supplement to the desired Staphylococcus test vial.
3. Place the remaining S2-SSA rehydrated supplement in the refrigerator.



620 Leshar Place, Lansing, MI 48912  
800/234-5333 (USA/Canada) or 517/372-9200 • fax: 517/372-2006  
e-mail: [foodsafety@neogen.com](mailto:foodsafety@neogen.com) • [www.neogen.com](http://www.neogen.com)

© Neogen Corporation, 2019. Neogen and Soleris are registered trademarks of Neogen Corporation.  
C0120352B

S2-SSA-0819