

ORAL FLUID MULTI-ANALYTE CALIBRATOR SET

CALIBRATOR SET INSTRUCTIONS PRODUCT #128100 (RTU)

FORENSIC USE ONLY

DESCRIPTION

Neogen® Corporation's Oral Fluid Multi-Analyte Calibrators consist of four, ready-to-use calibrators in a synthetic oral fluid matrix. This set is intended for the qualitative evaluation of NeoSal™ oral fluid samples when used with Neogen Corporation's Cocaine, Methamphetamine, Opiates, Phencyclidine (PCP) and THC Oral Fluid Kits. Please see the table below for a description of the individual analyte concentrations for each calibrator.

Oral Fluid Multi-Analyte Calibrators	S (+)				
	Benzoylecgonine	Methamphetamine	Morphine	PCP	(-) Δ9-THC
Negative Calibrator	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL
Low Positive Calibrator	2.5 ng/mL	6.25 ng/mL	5 ng/mL	1.25 ng/mL	0.5 ng/mL
Cut-off Calibrator	5 ng/mL	12.5 ng/mL	10 ng/mL	2.5 ng/mL	1 ng/mL
High Positive Calibrator	10 ng/mL	25 ng/mL	20 ng/mL	5 ng/mL	2 ng/mL

STORAGE AND STABILITY

The set is stable up to the printed expiration date, prior to opening and when stored at 4°C. Do not freeze.

MATERIALS PROVIDED

1. **Negative Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
2. **Low Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
3. **Cut-off Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
4. **High Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

ORAL FLUID MULTI-ANALYTE CALIBRATOR SET

CALIBRATOR SET INSTRUCTIONS PRODUCT #128100 (RTU)

FORENSIC USE ONLY

DESCRIPTION

Neogen® Corporation's Oral Fluid Multi-Analyte Calibrators consist of four, ready-to-use calibrators in a synthetic oral fluid matrix. This set is intended for the qualitative evaluation of NeoSal™ oral fluid samples when used with Neogen Corporation's Cocaine, Methamphetamine, Opiates, Phencyclidine (PCP) and THC Oral Fluid Kits. Please see the table below for a description of the individual analyte concentrations for each calibrator.

Oral Fluid Multi-Analyte Calibrators	S (+)				
	Benzoylecgonine	Methamphetamine	Morphine	PCP	(-) Δ9-THC
Negative Calibrator	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL
Low Positive Calibrator	2.5 ng/mL	6.25 ng/mL	5 ng/mL	1.25 ng/mL	0.5 ng/mL
Cut-off Calibrator	5 ng/mL	12.5 ng/mL	10 ng/mL	2.5 ng/mL	1 ng/mL
High Positive Calibrator	10 ng/mL	25 ng/mL	20 ng/mL	5 ng/mL	2 ng/mL

STORAGE AND STABILITY

The set is stable up to the printed expiration date, prior to opening and when stored at 4°C. Do not freeze.

MATERIALS PROVIDED

1. **Negative Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
2. **Low Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
3. **Cut-off Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
4. **High Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

ORAL FLUID MULTI-ANALYTE CALIBRATOR SET

CALIBRATOR SET INSTRUCTIONS PRODUCT #128100 (RTU)

FORENSIC USE ONLY

DESCRIPTION

Neogen® Corporation's Oral Fluid Multi-Analyte Calibrators consist of four, ready-to-use calibrators in a synthetic oral fluid matrix. This set is intended for the qualitative evaluation of NeoSal™ oral fluid samples when used with Neogen Corporation's Cocaine, Methamphetamine, Opiates, Phencyclidine (PCP) and THC Oral Fluid Kits. Please see the table below for a description of the individual analyte concentrations for each calibrator.

Oral Fluid Multi-Analyte Calibrators	S (+)				
	Benzoylecgonine	Methamphetamine	Morphine	PCP	(-) Δ9-THC
Negative Calibrator	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL
Low Positive Calibrator	2.5 ng/mL	6.25 ng/mL	5 ng/mL	1.25 ng/mL	0.5 ng/mL
Cut-off Calibrator	5 ng/mL	12.5 ng/mL	10 ng/mL	2.5 ng/mL	1 ng/mL
High Positive Calibrator	10 ng/mL	25 ng/mL	20 ng/mL	5 ng/mL	2 ng/mL

STORAGE AND STABILITY

The set is stable up to the printed expiration date, prior to opening and when stored at 4°C. Do not freeze.

MATERIALS PROVIDED

1. **Negative Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
2. **Low Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
3. **Cut-off Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
4. **High Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

ORAL FLUID MULTI-ANALYTE CALIBRATOR SET

CALIBRATOR SET INSTRUCTIONS PRODUCT #128100 (RTU)

FORENSIC USE ONLY

DESCRIPTION

Neogen® Corporation's Oral Fluid Multi-Analyte Calibrators consist of four, ready-to-use calibrators in a synthetic oral fluid matrix. This set is intended for the qualitative evaluation of NeoSal™ oral fluid samples when used with Neogen Corporation's Cocaine, Methamphetamine, Opiates, Phencyclidine (PCP) and THC Oral Fluid Kits. Please see the table below for a description of the individual analyte concentrations for each calibrator.

Oral Fluid Multi-Analyte Calibrators	S (+)				
	Benzoylecgonine	Methamphetamine	Morphine	PCP	(-) Δ9-THC
Negative Calibrator	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL	0 ng/mL
Low Positive Calibrator	2.5 ng/mL	6.25 ng/mL	5 ng/mL	1.25 ng/mL	0.5 ng/mL
Cut-off Calibrator	5 ng/mL	12.5 ng/mL	10 ng/mL	2.5 ng/mL	1 ng/mL
High Positive Calibrator	10 ng/mL	25 ng/mL	20 ng/mL	5 ng/mL	2 ng/mL

STORAGE AND STABILITY

The set is stable up to the printed expiration date, prior to opening and when stored at 4°C. Do not freeze.

MATERIALS PROVIDED

1. **Negative Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
2. **Low Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
3. **Cut-off Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.
4. **High Positive Calibrator:** 2.5 mL provided in a synthetic oral fluid matrix.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

TECHNICAL SUPPORT

For technical assistance, please contact our Technical Services Department at 859-254-1221 or email at techservice-toxicology@neogen.com. Representatives are available Monday–Friday from 8:00 am–6:00 pm EST.

COPYRIGHT

All rights reserved worldwide. No part of this publication may be reproduced, transmitted, transcribed, or stored in any information retrieval system, or translated into any human or computer language in any form or by any means (manual, electronic, mechanical, magnetic, optical, chemical, or otherwise) without expressed written permission.

WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the material from which its products are made are of standard quality. If any materials are defective, Neogen Corporation will provide a replacement product. Buyer assumes all risk and liability resulting from the use of this product and any of the predictive models. There is no warranty of merchantability of this product for any purpose. Neogen Corporation shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from use of this product.

Products labeled **FORENSIC USE ONLY** are tests that have not been FDA approved.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

©Neogen Corporation, 2015. Neogen is a registered trademark and Neosal is a trademark of Neogen Corporation, Lansing, MI. D128100 7/27/15

TECHNICAL SUPPORT

For technical assistance, please contact our Technical Services Department at 859-254-1221 or email at techservice-toxicology@neogen.com. Representatives are available Monday–Friday from 8:00 am–6:00 pm EST.

COPYRIGHT

All rights reserved worldwide. No part of this publication may be reproduced, transmitted, transcribed, or stored in any information retrieval system, or translated into any human or computer language in any form or by any means (manual, electronic, mechanical, magnetic, optical, chemical, or otherwise) without expressed written permission.

WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the material from which its products are made are of standard quality. If any materials are defective, Neogen Corporation will provide a replacement product. Buyer assumes all risk and liability resulting from the use of this product and any of the predictive models. There is no warranty of merchantability of this product for any purpose. Neogen Corporation shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from use of this product.

Products labeled **FORENSIC USE ONLY** are tests that have not been FDA approved.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

©Neogen Corporation, 2015. Neogen is a registered trademark and Neosal is a trademark of Neogen Corporation, Lansing, MI. D128100 7/27/15

TECHNICAL SUPPORT

For technical assistance, please contact our Technical Services Department at 859-254-1221 or email at techservice-toxicology@neogen.com. Representatives are available Monday–Friday from 8:00 am–6:00 pm EST.

COPYRIGHT

All rights reserved worldwide. No part of this publication may be reproduced, transmitted, transcribed, or stored in any information retrieval system, or translated into any human or computer language in any form or by any means (manual, electronic, mechanical, magnetic, optical, chemical, or otherwise) without expressed written permission.

WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the material from which its products are made are of standard quality. If any materials are defective, Neogen Corporation will provide a replacement product. Buyer assumes all risk and liability resulting from the use of this product and any of the predictive models. There is no warranty of merchantability of this product for any purpose. Neogen Corporation shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from use of this product.

Products labeled **FORENSIC USE ONLY** are tests that have not been FDA approved.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

©Neogen Corporation, 2015. Neogen is a registered trademark and Neosal is a trademark of Neogen Corporation, Lansing, MI. D128100 7/27/15

TECHNICAL SUPPORT

For technical assistance, please contact our Technical Services Department at 859-254-1221 or email at techservice-toxicology@neogen.com. Representatives are available Monday – Friday from 8:00 am–6:00 pm EST.

COPYRIGHT

All rights reserved worldwide. No part of this publication may be reproduced, transmitted, transcribed, or stored in any information retrieval system, or translated into any human or computer language in any form or by any means (manual, electronic, mechanical, magnetic, optical, chemical, or otherwise) without expressed written permission.

WARRANTY

Neogen Corporation makes no warranty of any kind, either expressed or implied, except that the material from which its products are made are of standard quality. If any materials are defective, Neogen Corporation will provide a replacement product. Buyer assumes all risk and liability resulting from the use of this product and any of the predictive models. There is no warranty of merchantability of this product for any purpose. Neogen Corporation shall not be liable for any damages, including special or consequential damage, or expense arising directly or indirectly from use of this product.

Products labeled **FORENSIC USE ONLY** are tests that have not been FDA approved.



944 Nandino Blvd., Lexington KY 40511 USA
800-477-8201 (USA/Canada) • 859-254-1221
Fax: 859/255-5532 • E-mail: inform@neogen.com • Web: www.neogen.com/Toxicology

©Neogen Corporation, 2015. Neogen is a registered trademark and Neosal is a trademark of Neogen Corporation, Lansing, MI. D128100 7/27/15